

Cite this: DOI: 10.1039/c0xx00000x

www.rsc.org/xxxxxx

ARTICLE TYPE

# A biosensor for the determination of high density lipoprotein cholesterol employing combined surfactant-derived selectivity and sensitivity enhancements

5 **Termeh Ahmadraji, Laura Gonzalez-Macia and Anthony J. Killard\****Received (in XXX, XXX) Xth XXXXXXXXXX 20XX, Accepted Xth XXXXXXXXXX 20XX*

DOI: 10.1039/b000000x

High density lipoprotein cholesterol (HDL-C) is a modifiable risk factor in cardiovascular disease and devices suitable for its determination at the point of care are critical to the future management of hypercholesterolaemia. An electrochemical biosensor for measuring HDL-C was developed. The biosensor was based on a homogeneous assay methodology for selective determination of HDL-C in combination with a printed electrochemical sensor for measuring the reduction of hydrogen peroxide at a silver paste electrode. The polyoxyethylene alkylene tribenzylphenyl ether surfactant (Emulgen B-66) was found to be capable of both the selective dissolution of HDL particles, as well as the enhanced electrocatalytic reduction of hydrogen peroxide. The resulting biosensor was shown to have a linear response to HDL-C from 0.5 to 4 mM ( $r^2=0.998$ ) with an average r.s.d. of 7%. The biosensor was also used to analyse HDL-C in thirteen serum samples and had good agreement with a commercial spectrophotometric precipitation-based assay ( $r=0.7222$ ;  $p < 0.058$ ).

## Introduction

20 Cholesterol levels, including high density lipoprotein cholesterol (HDL-C) are modifiable risk factors for cardiovascular disease (CVD), a condition which remains the number one global cause of death.<sup>1-3</sup> Levels of HDL-C above 60 mg/dL (1.55 mM) are considered to have a positive protective role in heart disease, while low HDL-C levels (less than 40 mg/dL or about 1 mM) are linked to an increase in heart attack risk. For this reason, the importance of measurement of HDL-C has been emphasized by the National Cholesterol Education Programme (NCEP) since the late 1980s.<sup>4, 5</sup> The measurement of HDL-C is also important for two other purposes:<sup>1, 6, 7</sup>

1. The calculation of low density lipoprotein cholesterol (LDL-C) using the Friedewald formula:  $LDL-C = Total\ cholesterol - (HDL-C + TG/5)$
2. The calculation of non-HDL cholesterol, determined by subtracting the HDL cholesterol concentration from the TC content. Non-HDL-C has been recommended as a target for preliminary CVD prevention.

The gold standard for measurement of HDL-C and other cholesterol is the method developed by the Centres for Disease Control and Prevention (CDC).<sup>7</sup> This method is highly complex and requires 5.0 mL of sample which is subjected to



ultracentrifugation, precipitation and measurement using the Abell-Kendall method. Since there are only a few laboratories capable of performing the ultracentrifugation steps necessary in the CDC method, the Cholesterol Reference Method Laboratory Network (CRMLN) developed a simpler method based on a modified dextran sulphate procedure.<sup>8</sup> However, this technique, like the CDC method, also required large sample volumes and also required multiple manual processing steps including the removal of triglycerides which still renders this method unsuitable in most clinical laboratories and in automated analysers.<sup>9, 10</sup>

In the past three decades, chemical precipitation methods, and more recently, homogeneous assays have been used to measure serum HDL-C in clinical laboratories. Homogeneous assays were a major step forward in improving the precision of earlier precipitation methods. Full automation eliminated manual pipetting, off-line pre-treatment, centrifugation and separation steps and improved assay precision, in line with recommended NCEP criteria. To date, there are several commercial colorimetric assays available for the quantitative measurement of HDL-C in serum.<sup>3, 11</sup> All of these determine the amount of  $H_2O_2$  produced from the enzymatic reaction of cholesterol present in HDL, to cholest-4-en-3-one, which can then be measured

spectrophotometrically.

Homogeneous assay methodologies have been developed which allow the direct and selective analysis of HDL-C in a single step.<sup>3</sup>

In one method, a polyoxyethylene alkylene tribenzylphenyl ether surfactant (Emulgen B-66) was found to be capable of the selective solubilisation of HDL-C, allowing the enzymatic reaction of HDL-C to H<sub>2</sub>O<sub>2</sub> as follows:



where ChEs is cholesterol esterase and ChOx is cholesterol oxidase. However, such assays are only appropriate for laboratory analyses. The movement of many routine blood tests away from the central laboratory to the point-of-care is a major trend in healthcare provision. A point of care device that measures HDL-C directly is very attractive in biomedical diagnostics and would be highly advantageous in the self-management of hypercholesterolemia.<sup>12-14</sup> Electrochemical techniques lend themselves well to the fabrication of low cost, point of care and disposable diagnostic devices. Thus, it would seem a common sense approach to develop electrochemical biosensor methodologies that are capable of measuring HDL-C using a similar principle, with the measurement of H<sub>2</sub>O<sub>2</sub> performed electrochemically.<sup>15, 16</sup> While a number of electrochemical biosensors for cholesterol have been developed,<sup>17-20, 20-30</sup> to date, there are just a few published examples of electrochemical biosensors for HDL-C.<sup>31, 32</sup> Kinoshita *et al.*, developed an amperometric sensor based on a homogeneous assay in which, a peroxidase-entrapped and ferrocene-embedded carbon paste electrode was used to measure the H<sub>2</sub>O<sub>2</sub> produced after enzymatic reaction of HDL-C. PEG-modified enzymes in the presence of  $\alpha$ -cyclodextrin sulphate and MgCl<sub>2</sub> were employed to impart selectivity to the measurement of HDL-C. This method was performed at 37°C and was only linear up to 0.04 mM. However, since it is important to measure HDL-C directly up to at least 2 mM preferentially at room temperature, their method would be unsuitable for the development of a point of care device. Foster *et al.*, developed an electrochemical device for HDL-C based on a precipitation methodology in which phosphotungstic acid (PTA) and MgCl<sub>2</sub> were employed as the precipitation reagents.<sup>32</sup> However, incorporation of this method on a disposable platform has also proved challenging.

In the present work, an electrochemical sensor capable of the reduction of H<sub>2</sub>O<sub>2</sub> at a modified screen-printed silver electrode was employed as the basis of a biosensor to perform the selective measurement of HDL-C. It has been shown previously that these electrodes, modified with lyotropic layers composed of surfactant and salt, exhibit the significantly enhanced electrocatalytic reduction of H<sub>2</sub>O<sub>2</sub>.<sup>15, 16</sup> This behaviour was exhibited for a broad range of surfactants including anionic, cationic and neutral types. Here, we demonstrate that the polyoxyethylene alkylene tribenzylphenyl ether surfactant Emulgen B-66, facilitates both the selective measurement of HDL-C in serum, as well as enhancing the electrocatalytic reduction of the H<sub>2</sub>O<sub>2</sub> at the

electrode when formed following the enzymatic catalysis of cholesterol esters and cholesterol with ChEs and ChOx, respectively. This resulted in a biosensor capable of the direct, room temperature measurement of HDL-C in the diagnostically relevant range of 0 to 4 mM.

## Experimental

### Materials

Dodecylbenzenesulphonic acid (DBSA-D0989) was purchased from TCI Europe. Polyoxyethylene octyl phenyl ether (Triton X-100), sodium chloride (NaCl), potassium chloride (KCl), potassium dihydrogen phosphate (KH<sub>2</sub>PO<sub>4</sub>), disodium hydrogen phosphate (Na<sub>2</sub>HPO<sub>4</sub>), *N,N*-dimethyl-*m*-toluidine and 4-aminoantipyrine (4-AAP) were purchased from Sigma-Aldrich (UK). 4-Aminoantipyrine HCl was from BDH (Dorset, UK). Cholesterol oxidase (O5F; 19.4 U/mg), cholesterol esterase (CE4F; 144 U/mg) and horseradish peroxidase (HRP, HRP4C; 295 U/mg) were purchased from BBI Enzymes (Gwent, UK). HDL-C and LDL-C isolated from human sera and dilipidated difibrinated serum (S139) were purchased from Scipac Ltd. (Kent, UK). 30% (v/v) hydrogen peroxide solution was purchased from Merck (Nottingham, UK). Polyoxyethylene alkylene tribenzylphenyl ether (Emulgen B-66) was kindly donated by Kao Corporation (Japan). The HDL and LDL/VLDL cholesterol assay kit was purchased from Abcam (Cambridge, UK). Silver (PF-410), carbon (Electrodag 6017SS), and Ag/AgCl (Electrodag 6038SS) screen printing inks were from Henkel (Netherlands).

### Methods

#### Assay optimisation

The optimisation of the concentration of Emulgen B-66 required for the selective determination of HDL-C over LDL-C was performed spectrophotometrically. Solutions of either 0.5 mM HDL-C or LDL-C in PBS were mixed with Triton X-100 or Emulgen B-66 in a microtitre plate with an assay mixture containing 39 U/mL ChEs, 23 U/mL ChOx, 0.075 % (w/v) 4-AAP, 14 U/mL HRP and 0.06 % (v/v) *N,N*-dimethyl-*m*-toluidine. The resulting absorbance was measured at 545 nm on a FLUOstar Optima plate reader (BMG Labtech, UK) with Optima software (version: 2.1) after incubation for three min. at room temperature.

#### Biosensor development

All electrochemical experiments were performed on a 3 x 3 mm screen printed silver paste electrode (SPE) with a Ag/AgCl screen printed reference electrode and carbon counter electrode. Electrodes were used in either an open stirred batch system in 4 mL 0.1 M PBS pH 6.8 or in a low volume thin layer cell of 8  $\mu$ L formed from a lid and 25  $\mu$ m spacer layer and referred to in the text as 'encapsulated' electrodes. All electrochemical measurements were carried out using a PGSTAT128N potentiostat with NOVA 1.6 software (Metrohm, UK). Electrodes were either used without further modification or modified with an inkjet-printed layer of DBSA/KCl as previously reported.<sup>15</sup> Inkjet printing was performed using a Dimatix Materials Printer DMP-2831 with Dimatix Drop manager DMP-2800 series software (Fujifilm Dimatix, Inc., US). The effect of a number of reagents on the electrocatalytic reduction of H<sub>2</sub>O<sub>2</sub> at the electrodes was


assessed in the presence of a range of assay reagents including 6% (v/v) Emulgen B-66, ChEs (39 U/mL), ChOx (23 U/mL) and HDL-C (1.5 mM) in serum, either individually or in combination. Measurement of HDL-C in serum was performed via the chronoamperometric measurement of the  $H_2O_2$  produced after reaction of the HDL-C with ChEs (39 U/mL) and ChOx (23 U/mL) in 6% Emulgen B-66 at -0.1 V vs Ag/AgCl at 420 s following 180 s incubation at room temperature.

## Results and discussion

### 10 Optimisation of sensor selectivity using Emulgen B-66

The polyoxyethylene alkylene tribenzylphenyl ether, Emulgen B-66 possesses a hydrophile-lipophile balance of 13.2 which is believed to result in the selective break down of HDL particles – as opposed to other lipoproteins such as LDL, (very low density cholesterol) VLDL and chylomicrons – thus allowing the selective enzymatic catalysis of HDL-bound cholesterol.<sup>33-35</sup> The exact mechanism of HDL solubilisation remains unclear. However, since apolipoprotein A-I is the major apolipoprotein in HDL, this surfactant may solubilize the polar lipids via a specific interaction with this apolipoprotein.<sup>35</sup> To assess the ability of Emulgen B-66 to achieve the selective break down of HDL, the recovery of cholesterol from serum samples containing HDL-C or LDL-C was measured spectrophotometrically.

25 Fig. 1(a) shows the percentage recovery of 0.5 mM LDL-C in serum in the presence of different concentrations of Emulgen B-66 relative to 0.5% (v/v) Triton X-100, which is a non-specific surfactant and results in the release of total cholesterol (TC).<sup>36,37</sup>




30 **Fig. 1.** Recovery efficiencies of (a) 0.5 mM LDL-C and (b) 0.5 mM HDL-C in serum to Emulgen B-66, relative to 0.5% (v/v) Triton X-100. Recovery in Triton X-100 was taken as 100%.

In this case, recovery of LDL-C was  $18\pm6\%$ ,  $12\pm5\%$  and  $2.5\pm2\%$  (v/v) for Emulgen B-66 concentrations of 1%, 2.5% and 6%, respectively. Conversely, the percentage recovery of 0.5 mM HDL-C in the presence of 6% Emulgen B-66 was found to be  $96\pm3\%$ , relative to 0.5% Triton X-100 (Fig. 1b).<sup>35</sup>

### 40 Effect of assay reagents on the reduction of hydrogen peroxide


In order to evaluate the electrochemical response of the electrode towards  $H_2O_2$  in the presence of the assay components necessary to selectively measure HDL-C, time-based amperometric measurement of  $H_2O_2$  was performed in the presence of individual or combined assay reagents. In the case of Emulgen B-66, enhanced reduction currents were observed in its presence (Fig. 2a). In the absence of surfactant, electrodes had a response of  $1.33\times 10^{-7}$  A (curve 1). However, in the presence of 6% (v/v) Emulgen B-66, a response of  $5.98\times 10^{-6}$  A was obtained for 3 mM  $H_2O_2$  (curve 3), which was a 39% current increase over that achieved for the previously reported<sup>16</sup> combination of DBSA and KCl of  $3.64\times 10^{-6}$  A (curve 2). This enhanced response might assist in achieving a lower limit of detection of  $H_2O_2$  with the sensor.



55 **Fig. 2.** (a) Amperometric measurement of 0.5 to 3 mM  $H_2O_2$  at -0.1 V vs. Ag/AgCl in 4 mL stirred batch solution using: curve 1) unmodified Ag SPEs in 0.1 M PBS pH 6.8 solution; curve 2) DBSA/KCl modified Ag SPEs in 0.1 M PBS 6.8 solution; curve 3) SPEs in 6% (v/v) Emulgen B-66 solution. (b) Cyclic voltammograms (scan rate of 0.1 V/s vs Ag/AgCl in 0.1 M PBS, pH 6.8) of: curve 1) unmodified Ag SPEs and; curve 2) electrodes measured after 3 h in 6% (v/v) Emulgen B-66.

As previously demonstrated, the electrocatalytic reduction of  $\text{H}_2\text{O}_2$  has been shown to be significantly enhanced at a screen printed, silver paste electrode modified with a lyotropic layer formed by surfactant and electrolyte.<sup>16</sup> The mechanism for this is not fully understood, but may be due to several effects including a change in silver paste morphology by creating a high surface area nanostructure, or the formation of micellar, hexagonal or lamellar structures by surfactant in the solution which become deposited onto the silver paste and creates an enhanced surface for the catalytic process.<sup>16, 38</sup> Emulgen B-66 is a nonionic surfactant which was prepared in an electrolyte solution of 0.1 M PBS pH 6.8. It is believed that an equivalent effect is achieved by this combination as has previously been demonstrated with other surfactant/salt combinations. However, further enhancement appears to result from operation of the sensor in a solution of this surfactant and electrolyte, as opposed to the modification of the surface alone with DBSA and KCl.<sup>16</sup> Fig. 2(b) shows the cyclic voltammograms for unmodified electrodes and those in 6% Emulgen B-66 in PBS. These again show significant modification of the surface with a capacitive double layer formed by Emulgen B-66 and NaCl, as has been previously observed.


The effect of the serum sample and the enzymes on the amperometric response of the sensor was also investigated using amperometry in 4 mL stirred solution. Fig. 3 shows the response of the modified sensor to  $\text{H}_2\text{O}_2$  in the presence of 6% Emulgen B-66 before, during and following exposure of the sensor to HDL-C, ChEs and ChOx. The response of the sensor was similar before (a) and after (b) the combined exposure of the electrode to these species. However, the response was reduced significantly in the presence of HDL-C (d) and ChOx (e) alone. The presence of HDL or ChOx appears to significantly disrupt formation of the lyotropic layer on the electrode surface. In the case of HDL, it has already been clearly demonstrated that it interacts selectively with Emulgen B-66 and may disrupt the lyotropic layers formed at the electrode surface.



**Fig. 3.** Amperometric response of sensors to  $\text{H}_2\text{O}_2$  (0.5 to 1.5 mM) in stirred solution containing 6% (v/v) Emulgen B-66: a) before exposure and; b) after exposure to a solution of ChEs (39 U/mL), ChOx (23 U/mL) and serum containing 1.5 mM HDL-C; c) in the presence of ChEs (39 U/mL); d) in the presence of serum containing 1.5 mM HDL-C; e) in the presence of ChOx (23 U/mL).

In the case of ChOx, it has also been found to be capable of disrupting phospholipid membranes via the “active site lid” mechanism.<sup>39</sup> Phospholipid membranes are also formed from the organisation of amphiphilic molecules and are analogous to the lyotropic phases formed by the interaction of Emulgen B-66 and electrolyte at the electrode surface. Disruption may also relate to the highly hydrophobic nature of the ChOx active site and the presence of additional hydrophobic domains on its surface.<sup>40</sup> The sensor response was not significantly affected by the presence of ChEs. These results also demonstrated the reversible nature of the effect of both lipoprotein and ChOx on the electrocatalytic response of the sensor, further suggesting that only the formed lyotropic phase was affected and not the underlying electrode structure.

The final response of the biosensor system to 8  $\mu\text{L}$   $\text{H}_2\text{O}_2$  in a thin layer cell is shown in Fig. 4. This response takes into account all processes which either enhance the electrocatalysis such as the presence of Emulgen B-66, or which interferes with it, such as HDL and ChOx. Measurement was again performed at -0.1 V vs Ag/AgCl and the current response recorded at 420 s. The sensor had excellent linearity from 0 to 10 mM  $\text{H}_2\text{O}_2$  ( $r^2=0.996$ ,  $n=3$ ), which makes it suitable for the direct determination of HDL-C in serum, based on a resulting one to one stoichiometric relationship between the concentration of cholesterol and the concentration of  $\text{H}_2\text{O}_2$  generated, assuming full enzymatic conversion.




**Fig. 4.** Amperometric response to 8  $\mu\text{L}$  of  $\text{H}_2\text{O}_2$  (-0.1 V vs Ag/AgCl at 420 s) in the presence of 6% (v/v) Emulgen B-66, 39 U/mL ChEs and 23 U/mL ChOx at, (slope= $3.85 \times 10^{-8}$  A/mM,  $r^2=0.996$ , dotted line=95% confidence interval,  $n=3$ ).


#### 75 Measurement of HDL-C

The encapsulated biosensor was first applied to the measurement of HDL-C in diluted serum and sensitivity and reproducibility studies were performed. The chronoamperometric responses of the biosensor in HDL-C from 0.5 to 4 mM are shown in Fig. 5. Of note is the potential step chronoamperometric response occurring as the generated  $\text{H}_2\text{O}_2$  is reduced at the electrode. In classical potential-step chronoamperometry, the signal decays to zero due to the total consumption of the reactant.<sup>41</sup> Theoretically, the amperometric response would fall to 93% after  $t = L^2/D$ , where  $L$  is the diffusion layer thickness (25  $\mu\text{m}$ ) and  $D$  is the diffusion coefficient of  $\text{H}_2\text{O}_2$ . Based on a value of  $D$  of  $1.71 \times 10^{-5}$

$5 \text{ cm}^2 \text{ s}^{-1}$ ,<sup>42</sup> this would occur after 365 ms. However, in this work, a pseudo steady-state response was evident after several hundred seconds. This is most likely due to a combination of barriers to diffusion slowing the process of complete reduction of the available  $\text{H}_2\text{O}_2$ <sup>32</sup> and the continued production of some  $\text{H}_2\text{O}_2$  via the enzymatic catalysis of cholesterol. Coulometry can also be employed as an alternative to amperometry. The amperometric response taken at 420 s was found to be proportional to the HDL-C concentration (Fig. 6). The biosensor had a linear response of  $4.49 \times 10^{-8} \text{ A/mM}$  ( $r^2 = 0.998$ ,  $n=3$ ) between 0.5 and 4 mM HDL-C with an average r.s.d. of 7.0 %.




**Fig. 5.** Amperometric responses of the developed biosensor to HDL-C in 0.1 M PBS, pH 6.8 at -0.1V (vs Ag/AgCl). Inset: Detail of response between 360 and 420 s. HDL-C concentrations: a) 0.5 mM, b) 1 mM, c) 2 mM, d) 3 mM, e) 4 mM.



**Fig. 6.** Biosensor response to HDL-C in dilapidated serum at -0.1 V (vs Ag/AgCl). a) Assay controls using dilapidated serum in the presence of assay components for 0 mM concentration and in the absence of ChEs and ChOx for 0.5 to 4 mM concentrations; b) Biosensor containing 6% (v/v) Emulgen B-66, 39 U/mL ChEs and 23 U/mL ChOx; (slope= $4.49 \times 10^{-8}$ ,  $r^2 = 0.998$ ,  $n=3$ )

The measurement of HDL-C in clinical serum samples was then

studied using the developed biosensor. The HDL-C concentration was also measured using a spectrophotometric precipitation assay methodology (Abcam, UK) and correlated against the developed biosensor (Fig. 7). The correlation had a slope of 0.85 with  $r=0.7222$  ( $p < 0.058$ ) for 13 samples with minimum measurement of two times per sample using sensor and just one time using assay kit due to limitations in sample availability (31 measurements in total). The Bland-Altman plot which calculates the mean difference between the two methods of measurement demonstrates no bias between the two methods across the measured range with most of the measurements within the 95% confidence limit (mean difference  $\pm 2\text{Sd}$ ).<sup>43</sup> The biosensor indicated slightly higher HDL-C concentrations in some measurements compared to the assay kit. Although the exact reason for that is not known, it may be due to varying levels of free cholesterol present in the samples, as this has been shown to affect the response of this type of assay.<sup>44</sup> It has also been reported that most homogeneous assay methodologies give positive predictive values due to the presence of intermediate density lipoproteins (IDL) in the serum, or in the presence of high levels of Lp(a).<sup>45</sup> While the biosensor was developed based on the homogeneous assay methodology, the Abcam assay kit is based on the precipitation principle.<sup>9, 10, 46, 47</sup> Therefore, discrepancies between the two methods are very likely to be due to the differences in the methods and how they process and respond to the complex mixtures of lipids and lipoproteins in the sample.



**Fig. 7.** Correlation of HDL-C in clinical serum samples measured by the biosensor and the Abcam assay kit (slope=0.85,  $r=0.72$ ). Inset: Bland-Altman plot of the difference between sensor/abcam assay kit against the mean measurement of two methods in 31 measurements. (dashed lines are mean difference  $\pm 2\text{Sd}$ )

## Conclusions

An electrochemical biosensor was developed for the selective measurement of HDL-C. The sensor was able to achieve both selectivity and sensitivity enhancements using the surfactant Emulgen B-66. The presence of this surfactant was shown to selectively dissolve HDL over LDL. In addition, it also resulted in the enhanced electrocatalysis of  $\text{H}_2\text{O}_2$  which is produced following the release of cholesterol ester and cholesterol from HDL and its catalysis by

ChEs and ChOx, respectively. The biosensor was shown to have good linearity across the diagnostically relevant range. The biosensor was successfully applied to the measurement of HDL-C in real samples, although deviation from other methods was observed in some of the samples which might be due to the presence of variable quantities of other lipids.

## Notes and references

Address: Centre for Research in Biosciences, Department of Biological, Biomedical and Analytical Sciences, University of the West of England, Coldharbour Lane, Bristol BS16 1QY, UK.

\* E-mail: [tony.killard@uwe.ac.uk](mailto:tony.killard@uwe.ac.uk)

† Electronic Supplementary Information (ESI) available: [details of any supplementary information available should be included here]. See DOI: 10.1039/b000000x/

‡ Footnotes should appear here. These might include comments relevant to but not central to the matter under discussion, limited experimental and spectral data, and crystallographic data.

1 Adult Treatment Panel III (ATP III), *Detection, Evaluation, and Treatment of High Blood Cholesterol in Adults*, National Cholesterol Education Program (NCEP), USA, 2002.

2 World Health Organization, *Global status report on noncommunicable diseases 2010*, World Health Organization, Geneva, 2011.

3 T. Ahmadraji and A. J. Killard, *Anal. Methods*, 2013, **3**, 3612-3625.

4 G. R. Warnick, *Clin. Chem.*, 1995, **41**, 1427.

5 M. D. Bethesda, *Clin Chem.*, 1995, **41**, 1414-1420.

6 J. H. Contois, G. R. Warnick and A. D. Sniderman, *J Clin Lipidol.*, 2011, **5**, 264-272.

7 N. Rifai, R. Warnick and M. H. Dominiczak, *Handbook of Lipoprotein Testing*, AACC Press, Washington, 2001.

8 M. M. Kimberly, E. T. Leary, T. G. Cole and P. P. Waymack, *Clin Chem*, 1999, **45**, 1803-1812.

9 E. M. Bairaktari, A. Elisaf, V. Katsaraki, A. D. Tsimihodimos, K. Tselepis, C. Siamopoulos and O. Tsolas, *Clinical Biochemistry*, 1999, **32**, 339-346.

10 N. Harris, V. Galpichian, J. Thomas, E. Iannotti, T. Law and N. Rifai, *Clin Chem*, 1997, **43**, 816-823.

11 S. Caudill, G. Cooper, S. Smith and G. Myers, *Clin. Chem.*, 1998, **44**, 1650-1658.

12 V. Gubala, L. F. Harris, A. J. Ricco and W. Tan MX D.E., *Anal. Chem.*, 2012, **84**, 487-515.

13 N. Rao, N. Johri, L. Robertson, D. Harvey, B. Persaud, D. P. Mikhailidis, M. Thomas and D. Nair, *Atherosclerosis*, 2009, **207**, 309-310 (DOI:10.1016/j.atherosclerosis.2009.07.013).

14 J. R. Taylor, *Ann. Pharmacother.*, 2004, **38**, 1252-1257.

15 L. Gonzalez-Macia, M. R. Smyth and A. J. Killard, *Electroanalysis*, 2012, **24**, 609-614 (DOI:10.1002/elan.201100444).

16 L. Gonzalez-Macia, *Electrochim. Acta*, 2011, **58**, 562-570.

17 S. S. J. Aravind, T. T. Baby, T. Arockiadoss, R. Rakhi and S. Ramaprabhu, *Thin Solid Films*, 2011, **519**, 5667-5672 (DOI:10.1016/j.tsf.2011.03.032).

18 S. K. Arya, M. Datta and B. D. Malhotra, *Biosensors and Bioelectronics*, 2008, **23**, 1083-1100 (DOI:10.1016/j.bios.2007.10.018).

19 R. Manjunatha, G. Shivappa Suresh, J. Savio Melo, S. F. D'Souza and T. Venkatarangaiah Venkatesha, *Talanta*, 2012, **99**, 302-309 (DOI:10.1016/j.talanta.2012.05.056).

20 Ö. Türkarlan, S. K. Kayahan and L. Toppare, *Sensors Actuators B: Chem.*, 2009, **136**, 484-488 (DOI:10.1016/j.snb.2008.10.016).

21 B. C. Özer, H. Özyörük, S. S. Çelebi and A. Yıldız, *Enzyme Microb. Technol.*, 2007, **40**, 262-265.

22 Y. J. Shin and J. Kameoka, *J IND ENG CHEM*, 2012, **18**, 193-197.

23 S. Singh, P. R. Solanki, M. K. Pandey and B. D. Malhotra, *Sensors Actuators B: Chem.*, 2006, **115**, 534-541 (DOI:<http://dx.doi.org/10.1016/j.snb.2005.10.025>).

24 S. Singh, A. Chaubey and B. D. Malhotra, *Anal. Chim. Acta*, 2004, **502**, 229-234 (DOI:10.1016/j.aca.2003.09.064).

25 S. Singh, R. Singhal and B. D. Malhotra, *Anal. Chim. Acta*, 2007, **582**, 335-343 (DOI:10.1016/j.aca.2006.09.010).

26 Y. Tsai, S. Chen and C. Lee, *Sensors Actuators B: Chem.*, 2008, **135**, 96-101 (DOI:10.1016/j.snb.2008.07.025).

27 C. Zhao, *Anal. Biochem.*, 2008, **383**, 25-30.

28 X. Tan, M. Li, P. Cai, L. Luo and X. Zou, *Anal. Biochem.*, 2005, **337**, 111-120 (DOI:10.1016/j.ab.2004.10.040).

29 X. Liu, Z. Nan, Y. Qiu, L. Zheng and X. Lu, *Electrochim. Acta*, 2013, **90**, 203-209 (DOI:<http://dx.doi.org/10.1016/j.electacta.2012.11.119>).

30 J. C. Vidal, E. García and J. R. Castillo, *Anal. Chim. Acta*, 1999, **385**, 213-222 (DOI:[http://dx.doi.org/10.1016/S0003-2670\(98\)00838-1](http://dx.doi.org/10.1016/S0003-2670(98)00838-1)).

31 H. Kinoshita, M. Torimura, K. Kano and T. Ikeda, *Ann. Clin. Biochem.*, 1998, **35**, 739-744.

32 R. Foster, J. Cassidy and E. O'Donoghue, *Electroanalysis*, 2000, **12**, 716-721 (DOI:10.1002/1521-4109(200005)12:9<716::AID-ELAN716>3.0.CO;2-S).

33 K. Hino, M. Nakamura and M. Manabe, *Method for quantitatively determining cholesterol*, C12Q1/44, JP, 1998.

34 H. Tamura, S. Nishino, T. Yamaguchi and K. Hino, *Test piece for measuring HDL-cholesterol*, C12Q1/44, JP, 2009.

35 M. Okada, H. Matsui, Y. Ito and A. Fujiwara, *J. Clin. Lab. Anal.*, 2001, **15**, 223-229.

- 
- 36 T. P. CARR, C. J. ANDRESEN and L. L. RUDEL, *Clin Biochem*, 1993, **26**, 39-42.
- 37 K. Yamamoto, R. Byme, C. Edelstein, B. Shen and A. M. Scanu, *J Lipid Res.*, 1984, **25**, 770-779.
- 5 38 R. J. Farn, *Chemistry and Technology of Surfactants*, Blackwell, Oxford, 2006.
- 39 K. B. Ghoshroy, W. Zhu and N. S. Sampson, *Biochem.*, 1997, **36**, 6133-6140.
- 40 A. Vrielink, L. F. Lloyd and D. M. Blow, *J. Mol. Biol.*, 1991, **219**,  
10 533-554.
- 41 P. T. Kissinger and W. R. Heineman, *Laboratory techniques in electroanalytical chemistry*, Marcel Dekker, Inc, 1996.
- 42 R. C. Peña, J. C. M. Gamboa, M. Bertotti and T. R. L. C. Paixão, *Int. J. Electrochem. Sci.*, 2011, **6**, 394-403.
- 15 43 J. M. Bland and D. G. Altman, *Stat. Methods Med. Res.*, 1999, **8**, 135-160.
- 44 S. Yamamoto, M. Yamamoto and K. Nakanishi, *Method of measuring lipid in specific lipoprotein.*, C12Q1/61, Japan, 2011.
- 45 M. Nakamura, Y. Kayamori, S. Sato and T. Shimamoto, in *Focus on*  
20 *cholesterol research*, ed. M. A. Kramer, Nova biomedical book, New York, 2006.
- 46 M. L. Arranz-Peña, *Clin. Chem.*, 1998, **44**, 2499.
- 47 C. Cobbaert, L. Zwang, F. Ceriotti, A. Modenese, P. Cremer, W. Herrmann, G. Hoss, J. Jarausch and M. Nauck, *Clinical Chemistry*, 1998,  
25 **44**, 779-789.